

Certificate of Analysis

Product: Rinse Solution
Product Number: RD-944071D
Shelf Life: 2 Years
Storage Conditions: 18 – 25 ° C
Quantity: 1 X 370 mL

Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
Sodium	120 ± 10	mmol/L
Potassium	10 ± 2	mmol/L
Chloride	110 ± 10	mmol/L
pH	7.0 ± 0.5	pH units

Details
Function: To rinse the sample flow path on the Radiometer 555 Analyzer.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing NaCl, KCl, surfactant and preservative. For in vitro diagnostic use only.

Certified by:



Kathy Cruz
Quality Assurance Manager
Diamond Diagnostics, Inc.