

## Certificate of Analysis

Product: ISE Cleaning Solution  
Product Number: OY-AUH1019D  
Shelf Life: 2 Years  
Storage Conditions: 18 – 25 ° C  
Quantity: 4 X 100 mL

Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
Sodium	1860 ± 100	mmol/L
pH	10 ± 1	pH units

Details
Function: For ISE cleaning procedure on Olympus Analyzers.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing 5.0% Sodium Hypochlorite. For in vitro diagnostic use only.

Certified by:



Kathy Cruz  
Quality Assurance Manager  
Diamond Diagnostics, Inc.