

Expected Ranges Chart

	pH			pCO ₂ mmHg			pO ₂ mmHg			Na ⁺ mmol/L			K ⁺ mmol/L			Ca ⁺⁺ mmol/L			Cl ⁻ mmol/L			Li ⁺ mmol/L		
	Mean	Min	Max	Mean	Min	Max	Mean	Min	Max	Mean	Min	Max	Mean	Min	Max	Mean	Min	Max	Mean	Min	Max	Mean	Min	Max
Blood Gas/ISE Analyzer																								
Diamond PROLYTE										138	134	142	4.44	4.24	4.64				95	91	99	1.12	1.05	1.19
Diamond SMARTLYTE, GEMLYTE										136	132	140	4.21	4.02	4.40	1.06	0.98	1.14	99	95	103	0.90	0.85	0.95
Diamond SMARTLYTE PLUS										137	133	141	4.46	4.26	4.66	1.10	1.02	1.18	99	95	104	1.01	0.95	1.07
Erba Mannheim_EC 90										153	149	158	4.58	4.37	4.78	1.09	1.01	1.17	113	108	118			
Eschweiler Combiline	7.325	7.295	- 7.354	37.7	35.2	- 40.1	166	154	- 177	145	141	- 149	4.46	4.26	- 4.66	1.11	1.02	- 1.19	106	101	- 111	1.04	0.98	- 1.11
Eschweiler Combisys II	7.325	7.295	- 7.354	34.4	32.1	- 36.6	166	155	- 178	145	141	- 149	4.46	4.26	- 4.66	1.11	1.02	- 1.19	106	101	- 111	1.04	0.98	- 1.11
Eschweiler ECOLYTE										143	139	- 147	4.46	4.26	- 4.66	1.11	1.02	- 1.19	106	101	- 111	1.04	0.98	- 1.11
Eschweiler ECOSYS II	7.325	7.295	- 7.354	34.4	32.1	- 36.6	166	155	- 178															
Horiba Yumizen E100										137	133	- 141	4.46	4.26	- 4.66	1.10	1.02	- 1.18	99	95	- 104	1.01	0.95	- 1.07
Medica EasyLyte Na/K, Na/K/Cl, Na/K/Li, Na/K/Cl/Li, Na/K/pH/Ca	7.334	7.304	- 7.363							140	136	- 145	4.18	3.99	- 4.36	1.17	1.08	- 1.25	99	94	- 103	1.18	1.11	- 1.25
Radiometer ABL 5	7.32	7.29	- 7.35	42.0	39.3	- 44.8	140	130	- 150															
Radiometer ABL 50, 500, 505, 510, 520	7.316	7.287	- 7.345	42.7	39.9	- 45.5	163	151	- 174	138	134	- 142	4.20	4.01	- 4.39	1.01	0.93	- 1.09						
Radiometer ABL 555	7.313	7.284	- 7.342	43.1	40.3	- 45.9	161	149	- 172	144	140	- 149	4.20	4.01	- 4.39	1.05	0.97	- 1.13						
Radiometer ABL 70, 77	7.318	7.289	- 7.347	48.3	45.1	- 51.4	136	126	- 145	143	139	- 147	4.30	4.10	- 4.49	1.02	0.95	- 1.10	101	96	- 105			
Radiometer ABL 600, 610, 620, EML-100	7.302	7.273	- 7.331	47.1	44.0	- 50.1	140	130	- 149	142	138	- 146	4.33	4.13	- 4.52	1.17	1.08	- 1.26	99	95	- 103			
Radiometer ABL 705, 710, 715, 720, 725	7.302	7.273	- 7.331	49.4	46.2	- 52.6	135	125	- 144	140	136	- 144	4.21	4.02	- 4.40	1.05	0.97	- 1.13	102	98	- 107			
Roche/AVL 990, 995	7.282	7.253	- 7.311	49.4	46.2	- 52.6	141	131	- 150	140	136	- 144	4.21	4.02	- 4.40	1.05	0.97	- 1.13	104	99	- 109			
Roche/AVL 9110, 9140	7.302	7.273	- 7.331							138	134	- 142	4.19	4.00	- 4.37	1.07	0.99	- 1.15						
Roche AVL 9120, 9130										140	136	- 144	4.11	3.93	- 4.30				109	104	- 114			
Roche/AVL 9180, 9181										136	132	- 140	4.21	4.02	- 4.40	1.06	0.98	- 1.14	99	95	- 103	0.90	0.85	- 0.95
Roche/AVL Cobas b 121	7.332	7.302	- 7.361	49.4	46.2	- 52.6	129	120	- 138	144	140	- 148	4.31	4.12	- 4.51	0.97	0.90	- 1.05	109	104	- 114			
Roche/AVL Cobas b 221	7.322	7.293	- 7.351	49.4	46.2	- 52.6	128	119	- 137	144	140	- 148	4.31	4.12	- 4.51	0.97	0.90	- 1.05	109	104	- 114			
Roche/AVL Compact Series	7.292	7.263	- 7.321	50.4	47.1	- 53.7	141	131	- 150															
Roche/AVL Cobas Mira ISE										142	138	- 146	4.25	4.05	- 4.44				108	103	- 113			
Siemens/Bayer 248	7.276	7.247	- 7.305	46.8	43.8	- 49.8	125	116	- 133															
Siemens/Bayer 348	7.367	7.338	- 7.396	47.9	44.8	- 51.0	124	115	- 133	141	136	- 145	4.36	4.17	- 4.56	1.13	1.04	- 1.21	101	97	- 106			
Siemens/Bayer 614, 634, 644, 654, 664	7.33	7.30	- 7.36							143	139	- 147	4.35	4.15	- 4.54	1.07	0.99	- 1.15	105	101	- 110	1.14	1.08	- 1.21
Siemens/Bayer 840, 845, 850, 855, 860, 865	7.341	7.311	- 7.370	45.5	42.5	- 48.4	140	131	- 150	136	132	- 140	4.19	4.01	- 4.38	1.00	0.92	- 1.07	99	94	- 103			
Siemens/Bayer RapidPoint 400, 405	7.326	7.296	- 7.355	49.4	46.2	- 52.6	139	129	- 148	138	134	- 142	4.21	4.02	- 4.40	1.05	0.97	- 1.13	106	102	- 111			

<p>IVD For In Vitro Diagnostic Use in vitro diagnostisch Usage In Vitro Para Uso Diagnostico In Vitro Utilizar Aparato En Diagnostico In Vitro In vitro diagnostisch 仅供体外诊断使用 для использования в лабораторных условиях</p>	<p>CE European Conformity CE-Konformitätskennzeichnung Conformité aux normes européennes Conformidad europea Conformidade com as normas europeias Europäisch übereinstimmend 符合欧洲 Европейская Адекватность</p>	<p>⚠ Temperature Limit Temperaturlimit Limite de température Limite de temperatura Temperaturgrenze 測温法温度限制 温度限制 Температурные ограничения</p>	<p>📖 Consult Instructions for Use Gebrauchsanweisung beachten Consulte le notice d'emploi Consulte las instrucciones de uso Consulte as instruções de utilização Bevor Inanspruchnahme 參閱說明書使用 Прочитайте инструкции по применению</p>	<p>LOT Lot Number Chargen-Nr. Número de lot Número de lote Batchnummer 批号 Номер серии</p>	<p>📅 Use by (YYYY-MM-DD) Verwendbar bis (JJJJ-MM-TT) Date de péremption (AAAA-MM-LJ) Usar hasta el (AAAA-MM-DD) Utilizar até (AAAA-MM-DD) Anwendbar bis (AAAA-MM-DD) Использовать для (год.месяц.день)</p>	<p>🏭 Manufactured by Hergestellt von Fabriqué par Fabricado por Fremstillet af ***製造 оказатель</p>	<p>EC REP Authorized Representative Bevollmächtigter Représentant agréé Representante autorizado Autorisiertes repräsentant Санкционированный представитель</p>	<p>REF Catalog Number Katalognummer Número de catálogo Número de catálogo Número de catálogo Каталог 产品编号 Номер каталога</p>
--	---	---	---	--	---	---	--	---

Delta 2304101 in red. Do not print

-0.049

0.8

2

2

0.08

-0.05

1

0.02