

Mission Control™
Blood Gas and Electrolyte Control - Level 4

LOT

2302156



2026/01

Expected Ranges Chart

Blood Gas/ISE Analyzer	pH			pCO ₂ mmHg			pO ₂ mmHg			Na ⁺ mmol/L			K ⁺ mmol/L			Ca ⁺⁺ mmol/L			Cl ⁻ mmol/L			
	Mean	Min	Max	Mean	Min	Max	Mean	Min	Max	Mean	Min	Max	Mean	Min	Max	Mean	Min	Max	Mean	Min	Max	
IL 1300 Series	7.280	7.251	- 7.309	55.6	51.9	- 59.2	321	303	- 339													
IL BGE	7.286	7.257	- 7.315	55.6	51.9	- 59.2	321	303	- 339													
Nova Stat Profile Systems	7.280	7.251	- 7.309	55.6	51.9	- 59.2	330	312	- 348	129	125	- 133	4.21	4.02	- 4.40	1.12	0.96	- 1.28	100	96	- 105	
Radiometer ABL 3, 30, 300, 330	7.289	7.260	- 7.318	54.6	51.0	- 58.1	299	283	- 316													
Radiometer ABL 4, 5	7.25	7.22	- 7.28	54.6	51.1	- 58.2	309	292	- 325													
Radiometer ABL 500, 510, 520	7.289	7.260	- 7.318	54.6	51.0	- 58.1	329	311	- 347													
Radiometer ABL 555	7.289	7.260	- 7.318	54.6	51.1	- 58.2	329	311	- 347	128	124	- 132	3.95	3.77	- 4.12	1.13	0.97	- 1.29	97	93	- 102	
Radiometer ABL 600, 610, 615, 620	7.289	7.260	- 7.318	54.6	51.1	- 58.2	329	311	- 347													
Roche/AVL 945, 947	7.264	7.235	- 7.293	54.6	51.0	- 58.1	320	302	- 338													
Roche/AVL 990, 995	7.264	7.235	- 7.293	54.6	51.0	- 58.1	320	302	- 338													
Roche/AVL Compact Series	7.264	7.235	- 7.293	54.6	51.0	- 58.1	320	302	- 338													
Siemens/Bayer 238	7.300	7.271	- 7.329	58.6	54.8	- 62.4	320	302	- 338													
Siemens/Bayer 248	7.296	7.267	- 7.325	54.6	51.0	- 58.1	308	291	- 325													
Siemens/Bayer 278, 280, 288	7.296	7.267	- 7.325	55.6	51.9	- 59.2	324	306	- 342	127	123	- 131	4.01	3.83	- 4.19	1.11	0.95	- 1.27	97	93	- 102	
Siemens/Bayer 348	7.265	7.236	- 7.294	55.7	52.1	- 59.3	296	280	- 312	133	129	- 137	4.05	3.87	- 4.23	1.11	0.95	- 1.27	97	93	- 102	
Siemens/Bayer 840, 845, 850, 855, 860, 865	7.294	7.265	- 7.323	54.6	51.1	- 58.2	282	267	- 298	125	121	- 129	4.02	3.84	- 4.20	1.02	0.88	- 1.17	89	85	- 93	

<p>For In Vitro Diagnostic Use In Vitro Diagnosticum Usage In Vitro Para Uso Diagnóstico In Vitro Utilizar Apenas Em Diagnóstico In Vitro Ti In Vitro diagnosticering 仅供体外诊断使用 Для использования в диагностике in-vitro</p>	<p>European Conformity CE-Konformitätskennzeichnung Conformité aux normes européennes Conformidad europea Conformidade com as normas europeias Europaisk overensstemmelse 符合欧 Европейская Адекватность</p>	<p>Temperature Limit Temperaturlimit Limite de température Limite de temperatura Limite de temperatura Temperaturgrænse 温度标准限制 Температурные ограничения</p>	<p>Consult Instructions for Use Gebrauchsanweisung beachten Consulter la notice d'emploi Consulte las instrucciones de uso Consulte as instruções de utilização Belyt brugsanvisninger 参考说明书使用 Рекомендации по применению</p>	<p>Lot Number Chargen-Nr. Número de lote Número de lote Batchnummer 批号 Номер серии</p>	<p>Use by (YYYY-MM-DD) Verwendbar bis (JJJJ-MM-TT) Date de péremption (AAAA-MM-JJ) Usar hasta el (AAAA-MM-GG) Utilizar até (AAAA-MM-DD) Anvend for (ÅÅÅÅ-MM-DD) 有效期至(YYYY-MM-DD) Используется для (год-месяц-день рождения) оизводитель</p>	<p>Manufactured by Hergestellt von Fabriqué par Fabricado por Fabricado por Fremstillet af ***製造 оизводитель</p>	<p>Authorized Representative Bevollmächtigter Représentant agréé Representante autorizado Representante autorizado Autoriseret representant 授权的代表 Санкционированный представитель</p>	<p>Catalog Number Katalognummer Número de catálogo Número de catálogo Número de catálogo Katalognr. 产品编号 Номер каталога</p>
---	--	--	---	--	---	--	---	---