

Certificate of Analysis

Product: 7.383 Buffer
Product Number: AV-BP0136D
Shelf Life: 2 Years
Storage Conditions: 18 – 25 ° C
Quantity: 3 X 90 mL


Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
pH	7.383 ± 0.005	pH units

Details
Function: To provide a calibration point for the pH electrode for AVL pH/Blood Gas Analyzers.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing KH ₂ PO ₄ , Na ₂ HPO ₄ , stabilizer and preservative. For in vitro diagnostic use only.

Traceability

Analyte	Standard	Test Method
pH	Secondary Standard, OEM Reagent Traceable to NIST	Siemens Blood Gas Analyzer (ISE)

Certified by:



Kathy Fisher
Director, Quality Assurance
Diamond Diagnostics Inc.