

Certificate of Analysis

Product: 6.841 Buffer
 Product Number: AV-BP0137D
 Shelf Life: 2 Years
 Storage Conditions: 18 – 25 ° C
 Quantity: 3 X 90 mL


Parameter	Specified Range	Units
Appearance pH	Clear, free of visible particulate matter 6.841 + 0.005	pH units

Details
Function: To provide a calibration point for the pH electrode for AVL pH/Blood Gas Analyzers.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing KH₂PO₄, Na₂HPO₄, stabilizer and preservative. For in vitro diagnostic use only.

Traceability

Analyte	Standard	Test Method
pH	NIST 2186i, 2186ii	Siemens Blood Gas Analyzer (ISE)

Certified by:



Kathy Fisher
 Director, Quality Assurance
 Diamond Diagnostics Inc.