

## Certificate of Analysis

Product: Electrode Conditioning Solution  
 Product Number: AV-BP0380D  
 Shelf Life: 4 Years  
 Storage Conditions: 18 – 25 ° C  
 Quantity: 1 X 100 mL


Parameter	Specified Range	Units
Appearance	<b>Clear, free of visible particulate matter</b>	
pH	<b>3.0 ± 1.0</b>	pH units

Details
Function: To condition the Na <sup>+</sup> electrode for Diamond SMARTLYTE, CARELYTE, GEMLYTE, Roche AVL 91xx, and 98x Electrolyte Analyzers.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing Ammonium hydrogen bifluoride.  For <i>in vitro</i> diagnostic use only.

### Traceability

Analyte	Standard	Test Method
pH	<b>Secondary Standard Traceable to NIST</b>	pH Meter

Certified by:



Kathy Fisher  
 Director, Quality Assurance  
 Diamond Diagnostics Inc.