

Certificate of Analysis

Product: ISE Cleaning Solution
 Product Number: AV-BP1025D
 Shelf Life: 4.5 Years
 Storage Conditions: 18 – 25 ° C
 Quantity: 1 X 100 mL


Parameter	Specified Range	Units
Appearance pH	Clear, free of visible particulate matter 11 ± 1	pH units

Details
Function: To clean the sample path on Diamond SMARTLYTE, CARELYTE, GEMLYTE, Roche AVL 91xx, 98x, Idexx VetLyte, and Alfa Wassermann StarLyte III Electrolyte Analyzers.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing Neodisher MA and NaCl. For <i>in vitro</i> diagnostic use only.

Traceability

Analyte	Standard	Test Method
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher
 Director, Quality Assurance
 Diamond Diagnostics Inc.