

Certificate of Analysis

Product: Cleaning Solution
 Product Number: AV-BP1889D
 Shelf Life: 3 Years
 Storage Conditions: 18 – 25 ° C
 Quantity: 3 X 90 mL


Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
Sodium	92 ± 5	mmol/L
pH	9 ± 1	pH units

Details
Function: For the daily cleaning of the AVL pH/BG systems for AVL pH/Blood Gas Analyzers.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing salts and surfactants. For in vitro diagnostic use only.

Traceability

Analyte	Standard	Test Method
Sodium	NIST 919b	IL 943 (Flame Photometry)
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher
 Director, Quality Assurance
 Diamond Diagnostics Inc.