

Certificate of Analysis

Product: BUN Reagent
Product Number: BK-443350D
Shelf Life: 1 Years
Storage Conditions: 2 – 8 ° C
Quantity: 1 X 500 mL


Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
Conductivity	~0.70	mS/cm
pH	~7.3	pH units

Details
Function: For the quantitative determination of urea nitrogen in serum or plasma on CX systems and for the determination of urea on CX Delta Systems.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing >100 kU/L urease (Jack Bean) and non-reactive ingredients necessary for optimum systems operation. For in vitro diagnostic use only.

Traceability

Analyte	Standard	Test Method
pH	Secondary Standard Traceable to NIST	pH Meter
Urease (Jack Bean)	Vendor COA	Beckman CX7 (Enzymatic)

Certified by:



Kathy Fisher
Director, Quality Assurance
Diamond Diagnostics Inc.