

Certificate of Analysis

Product: Glucose Reagent
 Product Number: BK-443355D
 Shelf Life: 1 Year
 Storage Conditions: 2 – 8 ° C
 Quantity: 1 X 500 mL


| Parameter | Specified Range | Units |
|------------------|--|----------|
| Appearance pH | Clear, free of visible particulate matter ~6.0 | pH units |

| Details |
|---|
| Function: For the quantitative determination of glucose in serum, plasma, urine and cerebrospinal fluid on the Beckman CX and CX Delta Systems. |
| Quality Specifications: Passed |
| Reagent Contents: An aqueous solution containing ≥ 350 kU/L glucose oxidase (A.niger), 1% ethanol, 40 mmol/L potassium iodide, preservatives and surfactants. For in vitro diagnostic use only. |

Traceability

| Analyte | Standard | Test Method |
|-----------------|---|-------------------------|
| pH | Secondary Standard Traceable to NIST | pH Meter |
| Glucose Oxidase | Vendor COA | Beckman CXT (Enzymatic) |

Certified by:



Kathy Fisher
 Director, Quality Assurance
 Diamond Diagnostics Inc.