

Certificate of Analysis

Product: Total Protein Reagent
Product Number: BK-450224D
Shelf Life: 2 Years
Storage Conditions: 18-25 °C
Quantity: 4 X 1000 mL


Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
Abs 545nm	0.05 $\pm$ 0.01	Absorbance units
pH	~12 - 13	pH units

Details
Function: For the quantitative determination of total protein in serum and CSF protein on the Beckman CX Synchron Delta.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing 8.81 mmol/ L Copper (II) Sulfate Pentahydrate, surfactant and preservative. For in vitro diagnostic use only.

Traceability

Analyte	Standard	Test Method
Abs 545nm	Secondary Standard, OEM Reagent	Beckman DLI 520 Spectrophotometer (Absorbance)
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher
Director, Quality Assurance
Diamond Diagnostics Inc.