

## Certificate of Analysis

Product: Calcium Reagent  
 Product Number: BK-450890D  
 Shelf Life: 2 Years  
 Storage Conditions: 18 – 25 °C  
 Quantity: 1 X 1000 mL


Parameter	Specified Range	Units
Appearance	<b>Clear, free of visible particulate matter</b>	
pH	<b>6.0 ± 1.5</b>	pH units

Details
Function: For the quantitative determination of Ca <sup>++</sup> in serum, plasma and urine on the Beckman CXSystems.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing 0.2 mmol/ L Aresenazo III, preservative, surfactants and other ingredients.  For in vitro diagnostic use only.

### Traceability

Analyte	Standard	Test Method
pH	<b>Secondary Standard Traceable to NIST</b>	pH Meter

Certified by:



Kathy Fisher  
 Director, Quality Assurance  
 Diamond Diagnostics Inc.