

## Certificate of Analysis


Product: Hitergent Solution  
 Product Number: BM-1135329D  
 Shelf Life: 3 Years  
 Storage Conditions: 18 – 25 °C  
 Quantity: 14 X 100 mL

Parameter	Specified Range	Units
Appearance	<b>Clear, free of visible particulate matter</b>	
pH	<b>~12.5</b>	pH units

Details
Function: To clean reaction cells and water bath.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing NaOH, nonionic surfactant, ethanolamine and preservative.  For in vitro diagnostic use only.

Traceability		
Analyte	Standard	Test Method
pH	<b>Secondary Standard Traceable to NIST</b>	pH Meter

Certified by:



Kathy Fisher  
 Director, Quality Assurance  
 Diamond Diagnostics Inc.