

Certificate of Analysis

Product: Hitergent Solution
Product Number: HT-1135329D
Shelf Life: 3 Years
Storage Conditions: 18 – 25° C
Quantity: 14 X 100 mL


Parameter	Specified Range	Units
Appearance pH	Clear, free of visible particulate matter ~12.5	pH units

Details
Function: To clean reaction cells and water bath.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing NaOH, nonionic surfactant, ethanolamine and preservative. For <i>in vitro</i> diagnostic use only.

Traceability

Analyte	Standard	Test Method
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher
Director, Quality Assurance
Diamond Diagnostics Inc.