

Certificate of Analysis

Product: Hitergent Solution
 Product Number: BM-409149D
 Shelf Life: 4 Years
 Storage Conditions: 18 – 25 ° C
 Quantity: 1 X 1000 mL


Parameter	Specified Range	Units
Appearance pH	Clear, free of visible particulate matter ~12.5	pH units

Details
Function: To clean reaction cells and water baths for Hitachi Series ISE Analyzers.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing NaOH, nonionic surfactant, ethanolamine and preservative. For in vitro diagnostic use only.

Traceability

Analyte	Standard	Test Method
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher
 Director, Quality Assurance
 Diamond Diagnostics Inc.