

## Certificate of Analysis

Product: Hitergent Solution  
Product Number: HT-409149D  
Shelf Life: 4 Years  
Storage Conditions: 18 – 25° C  
Quantity: 1 X 1000 mL


Parameter	Specified Range	Units
Appearance pH	<b>Clear, free of visible particulate matter</b> ~12.5	pH units

Details
<b>Function:</b> To clean reaction cells and water baths for Hitachi Series ISE Analyzers.
<b>Quality Specifications:</b> Passed
<b>Reagent Contents:</b> An aqueous solution containing NaOH, nonionic surfactant, ethanolamine and preservative.  <b>For <i>in vitro</i> diagnostic use only.</b>

### Traceability

Analyte	Standard	Test Method
pH	<b>Secondary Standard Traceable to NIST</b>	pH Meter

Certified by:



Kathy Fisher  
Director, Quality Assurance  
Diamond Diagnostics Inc.