

Certificate of Analysis

Product: ISE Diluent
 Product Number: BM-820637D
 Shelf Life: 4.5 Years
 Storage Conditions: 18 – 25 ° C
 Quantity: 1 X 1000 mL

Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
pH	5	pH units

Details
Function: To dilute samples for Na, K, Cl measurements on the ISE module. For use on Hitachi 704 and 902 Analyzers.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing Boric acid and preservative. For in vitro diagnostic use only.

Traceability

Analyte	Standard	Test Method
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher
 Director, Quality Assurance
 Diamond Diagnostics Inc.