

## Certificate of Analysis

Product: ISE Diluent  
Product Number: HT-820637D  
Shelf Life: 4.5 Years  
Storage Conditions: 18 – 25° C  
Quantity: 6 X 1000 mL


Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matters	
pH	5	pH units

Details
<b>Function:</b> To dilute samples for Na, K, Cl measurements on the ISE module. For use on Hitachi 704 and 902 Analyzers.
<b>Quality Specifications:</b> Passed
<b>Reagent Contents:</b> An aqueous solution containing Boric acid and preservative. <b>For <i>in vitro</i> diagnostic use only.</b>

### Traceability

Analyte	Standard	Test Method
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher  
Director, Quality Assurance  
Diamond Diagnostics Inc.