

Certificate of Analysis

Product: KCI Electrolyte Solution Product Number: HT-820639D Shelf Life: 4 Years Storage Conditions: 18 – 25° C Quantity: 1 X 1000 mL

Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
Potassium	1000 <u>+</u> 50	mmol/L

Details

Function: To provide constant electrical potential through the reference electrode on Hitachi 7xx, 914 and 917 Analyzers.

Quality Specifications: Passed

Reagent Contents: An aqueous solution containing KCl and preservative.

For in vitro diagnostic use only.

Traceability		
Analyte	Standard	Test Method
Potassium	NIST 918b	IL 943 (Flame Photometry)

Certified by: shi

Kathy Fisher Director, Quality Assurance Diamond Diagnostics Inc.

Diamond Diagnostics Inc. 333 Fiske Street Holliston, MA 01746-2048 USA Diamond Diagnostics Inc. Óradna utca 6 1044 Budapest HUNGARY Diamond Diagnostics India 14/3 Mathura Road 121003 Faridabad, Haryana INDIA **Diamond Diagnostics Ltda**

Av. Santo Antonio 7, Sala 102 CEP 42712-750 Lauro de Freitas - BA BRAZIL

ECO# 10514 SOP16-3047F Revision 11 Effective Date 02/28/25

diamonddiagnostics.com