

## Certificate of Analysis

Product: Wash Pack  
 Product Number: CD-104226D  
 Shelf Life: 2 Years  
 Storage Conditions: 18 – 25 ° C  
 Quantity: 1 X 450 mL


Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
Sodium	140 ± 3	mmol/L
Potassium	4.00 ± 0.1	mmol/L
Chloride	100 ± 3	mmol/L
Calcium	1.25 ± 0.1	mmol/L
pH	6.5 - 7.4	pH units

Details
Function: To wash the sample flow path on Ciba-Corning 348 Analyzers.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing salts, surfactant and preservative. For in vitro diagnostic use only.

### Traceability

Analyte	Standard	Test Method
Sodium	NIST 919b	IL 943 (Flame Photometry)
Potassium	NIST 918b	IL 943 (Flame Photometry)
Chloride	NIST 919b	Corning 925, SAT-500 Salt Analyzer (Titrimetric)
Calcium	NIST 915b	Roche AVL 9180
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher  
 Director, Quality Assurance  
 Diamond Diagnostics Inc.