

Certificate of Analysis

Product: Wash Pack
Product Number: SM-104226D
Shelf Life: 2 Years
Storage Conditions: 18 – 25° C
Quantity: 4 x 450 mL

Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
Sodium	140 $\pm$ 3	mmol/L
Potassium	4.00 $\pm$ 0.1	mmol/L
Chloride	100 $\pm$ 3	mmol/L
Calcium	1.25 $\pm$ 0.1	mmol/L
pH	6.5 - 7.4	pH units

Details

Function: To wash the sample flow path.

Quality Specifications: Passed


Reagent Contents: An aqueous solution containing salts, surfactant and preservative.

For *in vitro* diagnostic use only.

Traceability

Analyte	Standard	Test Method
Sodium	NIST 919b	IL 943 (Flame Photometry)
Potassium	NIST 918b	IL 943 (Flame Photometry)
Chloride	NIST 919b	Corning 925, SAT-500 Salt Analyzer (Titrimetric)
Calcium	NIST 915b	Roche AVL 9180
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher
Director, Quality Assurance
Diamond Diagnostics Inc.