

Certificate of Analysis

Product: Wash Pack
Product Number: CD-473497D
Shelf Life: 4 Years
Storage Conditions: 18 – 25 ° C
Quantity: 1 X 450 mL


Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
Sodium	155 $\pm$ 8	mmol/L
pH	6.8 – 7.4	pH units

Details
Function: To wash the sample flow path on the Ciba-Corning 248.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing 450 ml Wash Solution, Condition Kit and Deproteinizer. For in vitro diagnostic use only.

Traceability

Analyte	Standard	Test Method
Sodium	NIST 919b	IL 943 (Flame Photometry)
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher
Director, Quality Assurance
Diamond Diagnostics Inc.