

Certificate of Analysis

Product: Wash Pack Product Number: SM-473497D Shelf Life: 2 Years Storage Conditions: 18 – 25° C Quantity: 1 X 450 mL

Parameter	Specifie	ed Range	Units
Appearance	Clear, free of visible particulate matter		
Sodium	155	+ 8	mmol/L
рН	6.8	- 7.4	pH units

Details

Function: To wash the sample flow path.

Quality Specifications: Passed

Reagent Contents: An aqueous solution containing 450 ml Wash Solution, Condition Kit and Deproteinizer.

For *in vitro* diagnostic use only.

Traceability

Analyte	Standard	Test Method
Sodium	NIST 919b	IL 943 (Flame Photometry)
рН	Secondary Standard Tr	raceable to NIST pH Meter

Certified by:

+ Fishin

Kathy Fisher Director, Quality Assurance Diamond Diagnostics Inc.

Diamond Diagnostics Inc. 333 Fiske Street Holliston, MA 01746-2048 USA Diamond Diagnostics Inc. Óradna utca 6 1044 Budapest HUNGARY Diamond Diagnostics India 14/3 Mathura Road 121003 Faridabad, Haryana INDIA Diamond Diagnostics Ltda

Av. Santo Antonio 7, Sala 102 CEP 42712-750 Lauro de Freitas - BA BRAZIL

ECO# 10559 SOP16-3059F Revision 11 Effective Date 05/05/25

diamonddiagnostics.com