

Certificate of Analysis

Product: Conditioning Solution
 Product Number: CD-478701D
 Shelf Life: 3 Years
 Storage Conditions: 18 – 25 ° C
 Quantity: 1 X 2 mL


Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
pH	3 ± 1	pH units

Details
Function: To clean and condition the Na ⁺ and/or pH electrodes on the Ciba-Corning Analyzers.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing ≤ 1% of NH ₄ HF ₂ . For in vitro diagnostic use only.

Traceability

Analyte	Standard	Test Method
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher
 Director, Quality Assistance
 Diamond Diagnostics Inc.