

## Certificate of Analysis

Product: Release Solution  
 Product Number: CD-478866D  
 Shelf Life: 2 Years  
 Storage Conditions: 18 – 25 ° C  
 Quantity: 1 X 750 mL

Parameter	Specified Range	Units
Appearance	<b>Clear, free of visible particulate matter</b>	
pH	<b>&lt; 2</b>	pH units

Details
Function: To release CO2 from samples for tCO2 measurements for the Ciba-Corning 664 Analyzer.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing Lactic acid 50 mmol/L, surfactants and preservative.  For in vitro diagnostic use only.

**Traceability**

Analyte	Standard	Test Method
pH	<b>Secondary Standard Traceable to NIST</b>	pH Meter

Certified by:



Kathy Fisher  
 Director, Quality Assurance  
 Diamond Diagnostics Inc.