

Certificate of Analysis

Product: Reference Emulsion
Product Number: IL-09756904D
Shelf Life: 2 Years
Storage Conditions: 18 – 25 ° C
Quantity: 4 X 500 mL


Parameter	Specified Range	Units
Appearance	White opaque, free of visible particulate matter	
% T @ 405nm	9 - 13	
% T @ 660nm	45 - 55	

Details
Function: Reference solution for the IL 100-7000 ACL Analyzers.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing salt, silicone oil and emulsion. For in vitro diagnostic use only.

Traceability

Analyte	Standard	Test Method
%T@ 405nm	OEM Reagent	Beckman Spectrophotometer
%T@ 660nm	OEM Reagent	Beckman Spectrophotometer

Certified by:



Kathy Fisher
Director, Quality Assurance
Diamond Diagnostics Inc.