

## Certificate of Analysis

Product: Urine Diluent  
 Product Number: ME-2111D  
 Shelf Life: 3 Years  
 Storage Conditions: 18 – 25 ° C  
 Quantity: 1 X 500 mL

Parameter	Specified Range	Units
Appearance	<b>Clear, free of visible particulate matter</b>	
Conductivity	<b>0.830 ± .01</b>	mS/cm

Details
Function: To dilute the urine sample on the Diamond PROLYTE, Medica EasyLyte and IL Ilyte Electrolyte Analyzers.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing 5.2 mmol/L Magnesium Acetate. For <i>in vitro</i> diagnostic use only.

### Traceability

Analyte	Standard	Test Method
Conductivity	<b>Secondary Standard</b>	Total Dissolved Solids

Certified by:



Kathy Fisher  
 Director, Quality Assurance  
 Diamond Diagnostics Inc.