

## Certificate of Analysis

Product: Daily Rinse Kit  
 Product Number: ME-2118D  
 Shelf Life: 3 Years  
 Storage Conditions: 18 – 25 ° C  
 Quantity: 1 X 100 mL, 6 X 0.35g

Component	Description	Fill Level
ME-2118AD	Daily Rinse/Cleaning Diluent	100 mL
ME-2118BD	Pepsin	0.35g

**Daily Rinse:**

Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
pH	~ 1-3	pH units

**Details**

Function: To rinse and clean the sample path on the Diamond PROLYTE, Medica EasyLyte and IL Ilyte Electrolyte Analyzers.

Quality Specifications: Passed


Reagent Contents: An aqueous solution containing 1 bottle – 100 ml Daily Rinse/Cleaning Diluent, 6 bottles – 0.35 g pepsin/bottle.

For *in vitro* diagnostic use only.

**Traceability**

Analyte	Standard	Test Method
pH	Secondary Standard Traceable to NIST	pH Meter
Pepsin	Vendor COA	Enzyme Activity

Certified by:



Kathy Fisher  
 Director, Quality Assurance  
 Diamond Diagnostics Inc.