

Certificate of Analysis

Product: Flush Solution
 Product Number: ME-2309D
 Shelf Life: 3.5 Years
 Storage Conditions: 18 – 25 ° C
 Quantity: 1 X 50 mL


Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
pH	3 ± 1	pH units

Details
Function: To clean the fluid path on the Diamond PROLYTE, Medica EasyLyte and IL Ilyte Electrolyte Analyzers.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing 0.1 mol/L ammonium bifluoride. For <i>in vitro</i> diagnostic use only.

Traceability

Analyte	Standard	Test Method
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher
 Director, Quality Assurance
 Diamond Diagnostics, Inc.