

Certificate of Analysis

Product: Rinse Solution
Product Number: RD-943793D
Shelf Life: 2 Years
Storage Conditions: 18 – 25 ° C
Quantity: 1 X 340 mL


Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
Sodium	50.0 $\pm$ 3	mmol/L
Potassium	50.0 $\pm$ 3	mmol/L
Chloride	100 $\pm$ 5	mmol/L
pH	6.4 – 7.6	pH units

Details
Function: To provide a calibration point for the pH electrodes on the Radiometer 50, 500, 510, 520, 600, 610, and 620 Analyzers.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing NaCl, KCl, surfactant and preservative. For in vitro diagnostic use only.

Traceability

Analyte	Standard	Test Method
Sodium	NIST 919b	IL 943 (Flame Photometry)
Potassium	NIST 918b	IL 943 (Flame Photometry)
Chloride	NIST 919b	Corning 925, SAT-500 Salt Analyzer (Titrimetric)
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher
Director, Quality Assurance
Diamond Diagnostics Inc.