

Certificate of Analysis

Product: Cleaning Solution
 Product Number: RD-943795D
 Shelf Life: 3 Years
 Storage Conditions: 18 – 25 ° C
 Quantity: 1 X 140 mL

Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
pH	11 ± 1	pH units

Details
Function: to clean the sample path on the Radiometer 50, 500, 510, 520, 600, 610, and 620 Analyzers.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing 0.7% Deconex 51DR. For in vitro diagnostic use only.

Traceability

Analyte	Standard	Test Method
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher
 Director, Quality Assurance
 Diamond Diagnostics Inc.