

Certificate of Analysis

Product: Calibrator 2 Solution
 Product Number: RD-943839D
 Shelf Life: 2 Years
 Storage Conditions: 18 – 25 ° C
 Quantity: 1 X 140 mL


Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
Sodium	99.0 ± 10	mmol/L
pH	6.841 ± 0.005	pH units

Details
Function: To provide a calibration point for the pH electrode on the Radiometer 5 Analyzers.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing NaCl, NaOH, HEPES, surfactant and preservative. For in vitro diagnostic use only.

Traceability

Analyte	Standard	Test Method
Sodium	NIST 919b	IL 943 (Flame Photometry)
pH	NIST 2186i, 2186ii	Siemens Blood Gas Analyzer (ISE)

Certified by:



Kathy Fisher
 Director, Quality Assurance
 Diamond Diagnostics Inc.