

## Certificate of Analysis

Product: Cleaning Solution  
 Product Number: RD-943955D  
 Shelf Life: 3 Years  
 Storage Conditions: 18 – 25 ° C  
 Quantity: 1 X 140 mL


Parameter	Specified Range	Units
Appearance pH	<b>Clear, free of visible particulate matter</b> ~10 - 12	pH units

Details
Function: To clean lipid deposits from the sample flow path on the Radiometer 600, 610, 620, 625 & EML 100 Analyzers.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing 0.7% Deconex 51DR.  For in vitro diagnostic use only.

### Traceability

Analyte	Standard	Test Method
pH	<b>Secondary Standard Traceable to NIST</b>	pH Meter

Certified by:



Kathy Fisher  
 Director, Quality Assurance  
 Diamond Diagnostics Inc.