

## Certificate of Analysis

Product: Rinse Solution  
 Product Number: RD-943993D  
 Shelf Life: 3 Years  
 Storage Conditions: 18 – 25 ° C  
 Quantity: 1 X 340 mL


Parameter	Specified Range	Units
Appearance	<b>Clear, free of visible particulate matter</b>	
Sodium	<b>130.0 ± 5</b>	mmol/L
Potassium	<b>4.00 ± 0.1</b>	mmol/L
Chloride	<b>130 ± 5</b>	mmol/L
Calcium	<b>1.25 ± 0.1</b>	mmol/L
pH	<b>7 ± 0.6</b>	pH units

Details
Function: To rinse the measuring system after each calibration on the Radiometer 600, 610, 620, 625 & EML 100 Analyzers.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing Buffer, KCl, NaCl, CaC12, surfactant and preservative.  For in vitro diagnostic use only.

### Traceability

Analyte	Standard	Test Method
Sodium	<b>NIST 919b</b>	IL 943 (Flame Photometry)
Potassium	<b>NIST 918b</b>	IL 943 (Flame Photometry)
Chloride	<b>NIST 919b</b>	Corning 925, SAT-500 Salt Analyzer (Titrimetric)
Calcium	<b>NIST 915b</b>	Roche 9180 (ISE)
pH	<b>Secondary Standard Traceable to NIST</b>	pH Meter

Certified by:



Kathy Fisher  
 Director, Quality Assurance  
 Diamond Diagnostics Inc.