

Certificate of Analysis

Product: Rinse Solution
 Product Number: RD-944071D
 Shelf Life: 2 Years
 Storage Conditions: 18 – 25 ° C
 Quantity: 1 X 340 mL


Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
Sodium	120 ± 10	mmol/L
Potassium	10 ± 2	mmol/L
Chloride	110 ± 10	mmol/L
pH	7.0 ± 0.5	pH units

Details
Function: To rinse the sample flow path on the Radiometer 555 Analyzer.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing NaCl, KCl, surfactant and preservative. For in vitro diagnostic use only.

Traceability

Analyte	Standard	Test Method
Sodium	NIST 919b	IL 943 (Flame Photometry)
Potassium	NIST 918b	IL 943 (Flame Photometry)
Chloride	NIST 919b	Corning 925, SAT-500 Salt Analyzer (Titrimetric)
pH	Secondary Standard, OEM Reagent	Siemens Blood Gas Analyzer (ISE)

Certified by:



Kathy Fisher
 Director, Quality Assurance
 Diamond Diagnostics Inc.