

Certificate of Analysis

Product: Albumin Reagent
Product Number: BK-467858D
Shelf Life: 2 Years
Storage Conditions: 18 – 25 ° C
Quantity: 2 x 2000mL


Parameter	Specified Range	Units
Appearance	Amber color, free of visible particulate matter	
pH	5.20 ± 0.05	pH units

Details
Function: For the quantitative determination of Albumin in serum on the Olympus Beckman DxC and LX systems.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing 0.35mmol/L Bromocresol Purple, surfactant, and preservative. For in vitro diagnostic use only.

Traceability

Analyte	Standard	Test Method
pH	Secondary Standard Traceable to NIST	pH Meter (ISE)

Certified by:



Kathy Fisher
Director, Quality Assurance
Diamond Diagnostics Inc.