

Certificate of Analysis

Product: Phosphorus Reagent
 Product Number: BK-467868D
 Shelf Life: 2 Years
 Storage Conditions: 18-25 °C
 Quantity: 2x200mL Phosphorus
 Conc., 2x1800mL
 Phosphorus Diluent


Parameter	Specified Range	Units
Appearance	Clear Liquid	
pH	≤ 1	pH units

Details
Function: For the quantitative determination of phosphorus in serum or urine on the Olympus Beckman DxC and LX systems.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing 1.05 mmol/L Sulfuric Acid, 32.4 mmol/L Ammonium Para-Molybdate Tetrahydrate. For in vitro diagnostic use only.

Traceability

Analyte	Standard	Test Method
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher
 Director, Quality Assurance
 Diamond Diagnostics Inc.