

Certificate of Analysis

Product: ISE Reference
 Product Number: BK-467935D
 Shelf Life: 2 Years
 Storage Conditions: 18 – 25 ° C
 Quantity: 2 x 2000mL


Parameter	Specified Range	Units
Appearance pH	Blue Liquid 8.1 ± 0.1	pH units

Details
Function: For the quantitative determination of Na ⁺ , K ⁺ , Cl ⁻ , tCO ₂ , and Ca ⁺⁺ in serum or urine on the Olympus Beckman DxC and LX Systems.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing 45 mmol/L Tris Hydroxymethyl aminomethane, surfactant, and preservative. For in vitro diagnostic use only.

Traceability

Analyte	Standard	Test Method
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher
 Director, Quality Assurance
 Diamond Diagnostics Inc.