

Certificate of Analysis

Product: ISE Buffer
Product Number: BK-AUH1011D
Shelf Life: 3 Years
Storage Conditions: 18 – 25° C
Quantity: 4 X 2000 mL


Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
pH	7.82 ± 0.5	pH units

Details
Function: For the quantitative determination of Na ⁺ , K ⁺ , and Cl ⁻ concentrations in human serum and urine.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing 0.1 mol/L Triethanolamine and preservatives. For <i>in vitro</i> diagnostic use only.

Traceability

Analyte	Standard	Test Method
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher
Director, Quality Assurance
Diamond Diagnostics Inc.