

Certificate of Analysis

Product: ISE Reference Solution Product Number: BK-AUH1013D Shelf Life: 3 Years Storage Conditions: 18 – 25° C Quantity: 4 X 1000 mL

Parameter	Specified Range	Units	
Appearance	Clear, free of visible particulate matter		
Potassium	1000 <u>+</u> 50	mmol/L	

Details

Function: For the quantitative determination of Na+, K+, and Cl- concentrations in human serum and urine on Olympus Analyzers.

Quality Specifications: Passed

Reagent Contents: An aqueous solution containing 1.00 mol/L Potassium Chloride and preservatives.

For *in vitro* diagnostic use only.

Traceability			
Analyte	Standard	Test Method	
Potassium	NIST 918b	IL 943 (Flame Photometry)	

Certified by:

Fishin

Kathy Fisher Director, Quality Assurance Diamond Diagnostics Inc.

Diamond Diagnostics Inc. 333 Fiske Street Holliston, MA 01746-2048 USA Diamond Diagnostics Inc. Óradna utca 6 1044 Budapest HUNGARY Diamond Diagnostics India 14/3 Mathura Road 121003 Faridabad, Haryana INDIA **Diamond Diagnostics Ltda**

Av. Santo Antonio 7, Sala 102 CEP 42712-750 Lauro de Freitas - BA BRAZIL

diamonddiagnostics.com

ECO# 10526 SOP16-3213F Revision 07 Effective Date 04/04/25