

Certificate of Analysis

Product: ISE Cleaning Solution
Product Number: BK-AUH1019D
Shelf Life: 3 Years
Storage Conditions: 2 – 25° C
Quantity: 4 X 100 mL

Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
Hypochlorite	6.0 \pm 0.5	%
Sodium	1860 \pm 100	mmol/L
pH	12 \pm 2	pH units

Details

Function: For ISE cleaning procedure on Olympus Analyzers.

Quality Specifications: Passed


Reagent Contents: An aqueous solution containing 6.0% Sodium Hypochlorite. Recommended storage is refrigerated.

For *in vitro* diagnostic use only.

Traceability

Analyte	Standard	Test Method
Hypochlorite	OEM Reagent	Thiosulfate Titration
Sodium	NIST 918b	Flame Photometry
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher
Director, Quality Assurance
Diamond Diagnostics Inc.