

Certificate of Analysis

Product: ISE Cleaning Solution
 Product Number: OY-AUH1019D
 Shelf Life: 3 Years
 Storage Conditions: 2 – 25 ° C
 Quantity: 4 X 100 mL


Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
Hypochlorite	6.0 ± 0.5	%
Sodium	1860 ± 100	mmol/L
pH	10 ± 1	pH units

Details
Function: For ISE cleaning procedure on Olympus Analyzers.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing 6.0% Sodium Hypochlorite. <u>Recommended storage is refrigerated.</u> For in vitro diagnostic use only.

Traceability

Analyte	Standard	Test Method
Hypochlorite	OEM Reagent	Thiosulfate Titration
Sodium	NIST 918b	Flame Photometry
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher
 Director, Quality Assurance
 Diamond Diagnostics Inc.