

Certificate of Analysis

Product: Hct Level A
 Product Number: CD-570405D
 Shelf Life: 3 Years
 Storage Conditions: 18 – 25 ° C
 Quantity: 30 X 2 mL


Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
Sodium	65 ± 1	mmol/L
Potassium	4.00 ± 0.03	mmol/L
Chloride	59 ± 1	mmol/L
HCT (348 or eq)	25 ± 1	%
Conductivity	7.0 ± 0.6	mS/cm

Details
Function: For quality control of Hematocrit Analyzers.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing electrolytes and preservative. For in vitro diagnostic use only.

Traceability

Analyte	Standard	Test Method
Sodium	NIST 919b	IL 943 (Flame Photometry)
Potassium	NIST 918b	IL 943 (Flame Photometry)
Chloride	NIST 919b	Corning 925, SAT-500 Salt Analyzer (Titrimetric)
HCT	Secondary Standard, OEM Reagent	Siemens RapidLab 348 (Conductivity)

Certified by:



Kathy Fisher
 Director, Quality Assurance
 Diamond Diagnostics Inc.