

Certificate of Analysis

Product: BUN Reagent
 Product Number: BK-472482D
 Shelf Life: 1 Year
 Storage Conditions: 2-8°C
 Quantity: 2 x 200mL Urease
 Concentrate; 2 x 1800mL
 Urease Diluent

Component	Description	Fill Volume
BK-472482AD	Urease Diluent	1800 mL
BK-472482BD	Urease Concentrate	200 mL

Urease Diluent:

Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
pH	7.2 ± 0.1	pH Units

Urease Concentrate:

Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
pH	7.2 ± 0.1	pH Units

Details

Function: For the quantitative determination of urea nitrogen in serum or plasma on Olympus Beckman DxC and LX systems.

Quality Specifications: Passed


Reagent Contents: An aqueous solution containing 25 U/mL Jack Bean Urease and other non-reactive chemicals necessary for optimal system performance.

For in vitro diagnostic use only.

Traceability

Analyte	Standard	Test Method
pH	Secondary Standard Traceable to NIST	pH Meter
Urease	Secondary Standard, OEM Reagent	Beckman CX7 (Enzymatic)

Certified by:



Kathy Fisher
 Director, Quality Assurance