

## Certificate of Analysis

Product: Mission Trinity B, Level 1  
 Product Number: DD-96001  
 Shelf Life: 2 Years  
 Storage Conditions: 2 – 8°C  
 Quantity: 30 x 1.8 mL


Parameter	Specified Range	Units
Appearance	Purple Liquid	

Details
Function: For monitoring the performance of blood gas, electrolyte, metabolite, and CO-Oximeter (Bayer/Siemens) instrumentation for the analytes and analyzers listed on the Expected Values Chart.
Quality Specifications: Passed
<b>Reagent Contents: Buffered solution of electrolytes, glucose, lactate and dyes. It has been equilibrated with specific levels of CO<sub>2</sub>, O<sub>2</sub>, and N<sub>2</sub>. This control contains no human or biological materials.</b>
<b>For in vitro diagnostic use only.</b>

### Traceability

Analyte	Standard	Test Method
Sodium	NIST 919b	Bayer 8xx
Potassium	NIST 918b	Bayer 8xx
Chloride	NIST 919b	Bayer 8xx
Calcium	NIST 915b	Bayer 8xx
Glucose	NIST 917b	Cobas Mira Plus (Enzymatic)
Lactate	Pointe™ Lactate Std	Cobas Mira Plus (Enzymatic)

Certified by:



Kathy Fisher  
 Director, Quality Assurance  
 Diamond Diagnostics Inc.