

Certificate of Analysis

Product: Cleaning Solution
 Product Number: TB-20270479D
 Shelf Life: 2 Years
 Storage Conditions: 18 – 25 ° C
 Quantity: 1 X 25 mL


Parameter	Specified Range	Units
Appearance	Clear to slight yellow, free of visible particulate matter	
pH	7.5 ± 0.5	pH units

Details
Function: To clean the ISE sample port and sample path.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing enzymes, surfactants and preservatives. For in vitro diagnostic use only.

Traceability

Analyte	Standard	Test Method
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher
 Director, Quality Assurance
 Diamond Diagnostics Inc.