

Certificate of Analysis

Product: Urine Diluent
 Product Number: TB-20270482D
 Shelf Life: 3 Years
 Storage Conditions: 18 – 25 ° C
 Quantity: 500 mL


Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
pH	7.0 ± 0.5	pH units
Conductivity	1.2 ± 0.2	mS/cm

Details
Function: To dilute Urine samples on the Tokyo Boeki Biolis 24i and Prestige 24i.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing 3.8 mmol/L Magnesium Acetate. For <i>in vitro</i> diagnostic use only.

Traceability

Analyte	Standard	Test Method
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher
 Director, Quality Assurance
 Diamond Diagnostics Inc.