

## Certificate of Analysis

Product: Probe Wash 3  
 Product Number: SM-B01418351D  
 Shelf Life: 3 Year  
 Storage Conditions: 2 - 25°C  
 Quantity: 1 X 500 mL


Parameter	Specified Range	Units
Hypochlorite	5.5 - 6.5	%
Potassium	600 - 700	mmol/L
pH	13 - 14	pH Units

Details
<b>Function:</b> Function: For use as a probe cleaning agent on the Advia Chemistry Systems.
<b>Quality Specifications:</b> Passed
<b>Reagent Contents:</b> Contains 4.5% Potassium Hydroxide, 4.0% Sodium polyacrylic acid, 4.7% Sodium Hypochlorite, and surfactant. <u>Recommended storage is refrigerated.</u>  For in vitro diagnostic use only.

### Traceability

Analyte	Standard	Test Method
Hypochlorite	OEM Standard	Thiosulfate Titration
Potassium	NIST 918b	Flame Photometry
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher  
 Director, Quality Assurance  
 Diamond Diagnostics Inc.