

## Certificate of Analysis

Product: Cuvette Conditioner  
 Product Number: SM-B01417951D  
 Shelf Life: 2 Years  
 Storage Conditions: 18 - 25° C  
 Quantity: 1000 mL


Parameter	Specified Range	Units
Appearance	Clear to light yellow, free of visible particulate matter	
Potassium	13.0 ± 0.3	mmol/L
Conductivity	1.20 ± 0.05	mS/cm <sup>3</sup>
pH	5.0 ± 0.5	pH units

Details
Function: For use as a cuvette conditioner agent on the Advia Chemistry System.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing <0.5% K Sorbic Acid, <0.1% Citric Acid, and surfactant.  For in vitro diagnostic use only.

### Traceability

Analyte	Standard	Test Method
Potassium	NIST 918b	IL 943 (Flame Photometry)
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher  
 Director, Quality Assurance  
 Diamond Diagnostics Inc.